

UKCA Declaration of Conformity

Mahr GmbH
Carl-Mahr-Str. 1
D - 37073 Göttingen
Germany

Mahr UK Ltd
19 Drakes Mews, Crownhill, Milton Keynes
MK8 0ER
United Kingdom

We declare under our sole responsibility that the product

Name Digital Test Indicator

Type 800EWR

Item number 4306130, 4306140

From serial number 52040001

Is in conformity with the following standards

EN 55011:2016/ A1:2017/A11:2020
/A2:2021
EN IEC 61000-6-2:2019
EN IEC 61000-6-4:2019
EN IEC 63000:2018
EN 61010-
1:2010/A1:2019/A1:2019/AC:2019

Following the Directive(s)

2016 No 1091
2012 No 3032

Technical documentation team leader


Stefan Günther, Mahr GmbH

Executive director

Manuel Hüsken

Place and Date

Göttingen, 2025-04-15



Manuel Hüsken